

アゴニスト試料は、スポンサーより表10. の試料を提供された。

表10. アゴニスト試験で用いる試薬一覧

スポンサーID	状態	保存方法	受け入れ日	受け入れ者	Comments
H0009	Solid	Room Temp	4-Dec-08	Nakamura	
H0010	Solid	Room Temp	4-Dec-08	Nakamura	
H0011	Solid	Room Temp	4-Dec-08	Nakamura	
H0012	Solid	Room Temp	4-Dec-08	Nakamura	
H0013	Solid	Room Temp	4-Dec-08	Nakamura	
H0014	Solid	-4℃	4-Dec-08	Nakamura	
H0015	Solid	Room Temp	4-Dec-08	Nakamura	
H0016	Solid	Room Temp	4-Dec-08	Nakamura	

\* 標準及びコントロール物質は、前回購入したものを使用する。

アンタゴニスト試料は、スポンサーより表11. の試料を提供された。

表11. アンタゴニスト試験で用いる試薬一覧

スポンサーID	状態	保存方法	受け入れ日	受け入れ者	Comments
H00017	Liquid	Room Temp	4-Dec-08	Nakamura	
H00018	Solid	Room Temp	4-Dec-08	Nakamura	
H00019	Solid	Room Temp	4-Dec-08	Nakamura	
H00020	Solid	-4℃	4-Dec-08	Nakamura	
H00021	Solid	Room Temp	4-Dec-08	Nakamura	
H00022	Solid	-4℃	4-Dec-08	Nakamura	
H00023	Solid	-4℃	4-Dec-08	Nakamura	
H00024	Solid	Room Temp	4-Dec-08	Nakamura	

\* 標準及びコントロール物質は、前回購入したものを使用する。

### 3) 結果

Lumi-cell ER 細胞の 3) - 1 Substance solubility, 3) - 2 Range finding testing, 3) - 3 Comprehensive testing のアゴニスト活性及びアンタゴニスト活性についての EC<sub>50</sub> もしくは、IC<sub>50</sub> 及び Dose response curve の経歴データベースを作成した。

3) - 1 Substance solubility の結果は、下記の通りであった。

アゴニスト試料

- H0009 1mg/ml in DMSO, 0.01mg/ml in DMSO/aqueous cell culture media

- ・ H0010 1mg/ml in DMSO, 0.01mg/ml in DMSO/aqueous cell culture media
- ・ H0011 1mg/ml in DMSO, 0.01mg/ml in DMSO/aqueous cell culture media
- ・ H0012 10mg/ml in DMSO, 0.1mg/ml in DMSO/aqueous cell culture media
- ・ H0013 10mg/ml in DMSO, 0.1mg/ml in DMSO/aqueous cell culture media
- ・ H0014 10mg/ml in DMSO, 0.1mg/ml in DMSO/aqueous cell culture media
- ・ H0015 10mg/ml in DMSO, 0.1mg/ml in DMSO/aqueous cell culture media
- ・ H0016 10mg/ml in DMSO, 0.1mg/ml in DMSO/aqueous cell culture media

アンタゴニスト試料

- ・ H0017 2mg/ml in DMSO, 0.02mg/ml in DMSO/aqueous cell culture media
- ・ H0018 2mg/ml in DMSO, 0.02mg/ml in DMSO/aqueous cell culture media
- ・ H0019 2mg/ml in DMSO, 0.02mg/ml in DMSO/aqueous cell culture media
- ・ H0020 2mg/ml in DMSO, 0.02mg/ml in DMSO/aqueous cell culture media
- ・ H0021 20mg/ml in DMSO, 0.2mg/ml in DMSO/aqueous cell culture media
- ・ H0022 20mg/ml in DMSO, 0.2mg/ml in DMSO/aqueous cell culture media
- ・ H0023 2mg/ml in DMSO, 0.02mg/ml in DMSO/aqueous cell culture media
- ・ H0024 20mg/ml in DMSO, 0.2mg/ml in DMSO/aqueous cell culture media


3) - 2 Range finding testingの結果は、表12. 及び表13. 図11. 図12. の通りであった。

表12. アゴニスト試料における Range finding testing

Experiments: Phase IIb Range Finder Testing					
試験 I.D.	試料 Code	日付	Plate Induction	採用の有無	Rationale for Unacceptability
AgRF 1	H0009	23 Sep. 08	8.3	Used	Acceptable
	H0010	23 Sep. 08		Used	Acceptable
	H0011	23 Sep. 08		Used	Acceptable
	H0012	23 Sep. 08		Used	Acceptable
AgRF 2	H0013	23 Sep. 08	5.9	Used	Acceptable
	H0014	23 Sep. 08		Used	Acceptable
	H0015	23 Sep. 08		Used	Acceptable
	H0016	23 Sep. 08		Used	Acceptable


表 1 3. アゴニスト試料における Range finding testing

Experiments: Phase IIb Range Finder Testing					
試験 I.D.	試料 Code	日付	Plate Induction	採用の有無	Rationale for Unacceptability
AntRF 1	H0017	23 Sep. 08	8.7	Used	Acceptable
	H0018	23 Sep. 08		Used	Acceptable
	H0019	23 Sep. 08		Used	Acceptable
	H0020	23 Sep. 08		Used	Acceptable
AntRF 2	H0021	23 Sep. 08	8.4	Used	Acceptable
	H0022	23 Sep. 08		Used	Acceptable
	H0023	23 Sep. 08		Used	Acceptable
	H0024	23 Sep. 08		Used	Acceptable




○ : Starting Concentration used for Comprehensive testing  
 H0009, H0010, H0011, H0012 is Half-Log serial dilution

図 1 1 - 1. Range finder testing (アゴニスト試験) のグラフ




○ : Starting Concentration used for Comprehensive testing  
 H0013, H0015, H0016 is Double serial dilution  
 H0014 is Half-Log serial dilution

図 1 1 - 2. Range finder testing (アゴニスト試験) のグラフ



○ : Starting Concentration used for Comprehensive testing  
 H0017~H0020 is Double serial dilution

図 1 2 - 1. Range finding testing (アンタゴニスト試験) のグラフ



○ : Starting Concentration used for Comprehensive testing  
 H0021~H0024 is Double serial dilution

図 1 2 - 2. Range finding testing (アンタゴニスト試験) のグラフ

3-3 Comprehensive testing の結果は、表 1 4. 及び表 1 5. 図 1 3. 図 1 4. の通りであった。

表 1 4. Comprehensive testing (アンタゴニスト試験) の結果一覧

Experiments: Phase IIb Comprehensive Testing: Agonist						
試験 I.D.	試料 Code	日付	Plate Reduction	IC <sub>50</sub> (µg/mL)	採用の有無	不採用の理由
CT 1-1	H0009	07-Nov-08	3.5	0.83	Used	Acceptable
	H0010			0.27	Used	Acceptable
CT 2-1	H0011	07-Nov-08	1.6	1.7x10 <sup>-6</sup>	Repeated	Fail <sup>1</sup>
	H0012			NA	Repeated	Fail <sup>1</sup>
CT 3-1	H0013	07-Nov-08	4.1	1.1x10 <sup>+1</sup>	Used	Acceptable
	H0014			0.051 <sup>2</sup>	Used	Acceptable
CT 4-1	H0015	07-Nov-08	2.2	21	Repeated	Fail <sup>2</sup>
	H0016			0.80	Repeated	Fail <sup>2</sup>
CT 1-2	H0009	21-Nov-08	4.3	0.99	Used	Acceptable
	H0010			0.26	Used	Acceptable
CT 2-2	H0011	21-Nov-08	4.9	3.0x10 <sup>-6</sup>	Used	Acceptable
	H0012			0.49 <sup>2</sup>	Used	Acceptable
CT 3-2	H0013	21-Nov-08	4.5	1.9x10 <sup>+1</sup>	Repeated	Fail <sup>3</sup>
	H0014			0.12	Repeated	Fail <sup>3</sup>
CT 4-2	H0015	21-Nov-08	4.2	Positive <sup>4</sup>	Used	Acceptable
	H0016			1.4 <sup>2</sup>	Used	Acceptable
CT 1-3	H0009	04-Dec-08	3.1	0.43	Used	Acceptable
	H0010			0.21	Used	Acceptable
CT 2-3	H0011	04-Dec-08	3.1	1.9x10 <sup>-6</sup>	Used	Acceptable
	H0012			0.79	Used	Acceptable
CT 3-3	H0013	04-Dec-08	3.0	1.5	Repeated	Fail <sup>5</sup>
	H0014			0.0050	Repeated	Fail <sup>5</sup>
CT 4-3	H0015	04-Dec-08	3.0	Negative	Used	Acceptable
	H0016			1.3	Used	Acceptable
CT 2-4	H0011	19-Dec-08	4.7	2.6x10 <sup>-6</sup>	Used	Acceptable
	H0012			0.41	Used	Acceptable
CT 3-4	H0013	19-Dec-08	4.8	1.9x10 <sup>+1</sup>	Used	Acceptable
	H0014			0.16	Used	Acceptable

Experiments: Phase IIb Comprehensive Testing: Agonist						
試験 I.D.	試料 Code	日付	Plate Reduction	IC <sub>50</sub> (µg/mL)	採用の有無	不採用の理由
CT 4-4	H0015	19-Dec-08	3.3	Negative	Used	Acceptable
	H0016			8.2 <sup>2</sup>	Used	Acceptable
CT 3-5	H0013	25-Dec-08	3.3	2.5x10 <sup>+1</sup>	Used	Acceptable
	H0014			0.14 <sup>2</sup>	Used	Acceptable

NA = Not applicable, EC<sub>50</sub> values are not calculated for substances tested on plates not meeting acceptance criteria.

<sup>1</sup>The reason for the failure is because 1) DMSO Control / E2 Reference Standard EC50/ Methoxchlor is higher than the criteria and 2) Induction is lower than the criteria.

<sup>2</sup> The reason for the failure is because 1) DMSO Control / Methoxchlor is higher than the criteria and 2) Induction is lower than the criteria

<sup>3</sup> The reason for the failure is because E2 Reference Standard curve did not show sigmoid shape.

<sup>4</sup> Positive for agonism at 25 x 10<sup>0</sup> µg/mL

<sup>5</sup>The reason for the failure is because Methoxchlor Control is higher than the criteria.

表 1 5 . Comprehensive testing (アンタゴニスト試験) の結果一覧

Experiments: Phase IIb Comprehensive Testing: Antagonist						
試験 I.D.	試料 Code	日付	Plate Reduction	IC <sub>50</sub> (µg/mL)	採用の有無	不採用の理由
CT 1-1	H0017	14-Nov-08	9.8	positive <sup>1</sup>	Used	Acceptable
	H0018			positive <sup>2</sup>	Used	Acceptable
CT 2-1	H0019	14-Nov-08	8.8	negative	Used	Acceptable
	H0020			negative	Used	Acceptable
CT 3-1	H0021	14-Nov-08	13	550	Used	Acceptable
	H0022			25	Used	Acceptable
CT 4-1	H0023	14-Nov-08	13	negative	Used	Acceptable
	H0024			positive <sup>3</sup>	Used	Acceptable

Experiments: Phase IIb Comprehensive Testing:Antagonist						
試験 I.D.	試料 Code	日付	Plate Reduction	IC <sub>50</sub> (μg/mL)	採用の有無	不採用の理由
CT 1-2	H0017	29-Nov-08	8.2	negative	Repeated	Fail <sup>4</sup>
	H0018			negative	Repeated	Fail <sup>4</sup>
CT 2-2	H0019	29-Nov-08	10	negative	Repeated	Fail <sup>4</sup>
	H0020			positive <sup>5</sup>	Repeated	Fail <sup>4</sup>
CT 3-2	H0021	29-Nov-08	9.7	28	Used	Acceptable
	H0022			26	Used	Acceptable
CT 4-2	H0023	29-Nov-08	12	negative	Used	Acceptable
	H0024			positive <sup>2</sup>	Used	Acceptable
CT 1-3	H0017	12-Dec-08	10	negative	Used	Acceptable
	H0018			negative	Used	Acceptable
CT 2-3	H0019	12-Dec-08	11	negative	Used	Acceptable
	H0020			positive <sup>5</sup>	Used	Acceptable
CT 3-3	H0021	12-Dec-08	9.6	18	Used	Acceptable
	H0022			27	Used	Acceptable
CT 4-3	H0023	12-Dec-08	12	negative	Used	Acceptable
	H0024			positive <sup>1</sup>	Used	Acceptable
CT 1-4	H0017	19-Dec-08	8.4	positive <sup>1</sup>	Used	Acceptable
	H0018			positive <sup>6</sup>	Used	Acceptable
CT 2-4	H0019	19-Dec-08	8.8	negative	Used	Acceptable
	H0020			positive <sup>5</sup>	Used	Acceptable

<sup>1</sup> positive for antagonism at 1.0\*10+1μg/mL & 5.0\*10+0μg/mL

<sup>2</sup> positive for antagonism at 1.0\*10+1μg/mL & 5.0\*10+0μg/mL & 2.5\*10+0μg/mL


<sup>3</sup> positive for antagonism at 1.0\*10+1μg/mL & 5.0\*10+0μg/mL & 2.5\*10+0μg/mL & 1.25\*10+0μg/ml & 6.3\*10-1μg/ml

<sup>4</sup> for the failure is because Raloxifene/E2 Reference Standard IC50 is lower than the criteria.

<sup>5</sup> positive for antagonism at 1.0\*10+1μg/mL


<sup>6</sup> positive for antagonism at 5.0\*10+0μg/mL






\* Line resents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図 1 3 - 1 . Comprehensive testing (アゴニスト試験) H0009 グラフ




\* Line resents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図 1 3 - 2 . Comprehensive testing (アゴニスト試験) H0010 グラフ



\* Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean


図13-3. Comprehensive Testing (アゴニスト試験) H0011 グラフ



○: Deleted 1 high concentration points to calculate  $EC_{50}$  of H0012.


<sup>1</sup> Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図13-4. Comprehensive Testing (アゴニスト試験) H0012 グラフ



\* Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean


図 1 3 - 5 . Comprehensive testing (アゴニスト試験) H0013 グラフ



○ : Deleted 1 high concentration points to calculate EC<sub>50</sub> of H0014.


<sup>1</sup> Line represents the mean of three E2 replicates plus three times the standard deviation of the mean of E2.

図 1 3 - 6 . Comprehensive testing (アゴニスト試験) H0014 グラフ



\* Line resents the mean of three E2 replicates plus three times the standard deviation of the E2 mean


図 1 3 - 7 . Comprehensive testing (アゴニスト試験) H0015 グラフ



○ : Deleted 3 high concentration points to calculate EC<sub>50</sub> of H0016


<sup>1</sup> Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図 1 3 - 8 . Comprehensive testing (アゴニスト試験) H0016 グラフ




\* Lire represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-1. Comprehensive testing (アンタゴニスト試験) H0017 グラフ




<sup>1</sup> Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-2. Comprehensive testing (アンタゴニスト試験) H0018 グラフ




\* Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-3. Comprehensive testing (アンタゴニスト試験) H0019 グラフ




\* Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-4. Comprehensive testing (アンタゴニスト試験) H0020 グラフ




\* Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-5. Comprehensive testing (アンタゴニスト試験) H0021 グラフ




\* Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-6. Comprehensive testing (アンタゴニスト試験) H0022 グラフ



\* Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-7. Comprehensive testing (アンタゴニスト試験) H0023 グラフ



<sup>1</sup> Line represents the mean of three E2 replicates plus three times the standard deviation of the E2 mean

図14-8. Comprehensive testing (アンタゴニスト試験) H0024 グラフ



#### 4) 考察

アゴニスト試験において、16 試験中 4 試験が Criteria に入らなかったため、Fail 判定となった。アンタゴニスト試験において、14 試験中 2 試験が Criteria に入らなかったため、Fail 判定となった。表 9 に Phase II b の Criteria を示している。主な原因は、DMSO Control 及び一部の well における過剰な活性を引き起こした為の Criteria 判定以上の数値を示すことに原因があり、Agonist 試験であれば、DMSO Control 過剰に伴い、E<sub>2</sub>Reference Standard EC<sub>50</sub> / Induction の Fail 判定や Methoxchlor Control 過剰の Fail 判定、アンタゴニスト試験であれば、Ral/E<sub>2</sub> Reference Standard IC<sub>50</sub> の Fail 判定が起こっている。培養細胞の継代数が 50 代に至っていないものなるべく早期の細胞を使用することが必要かもしれない。Criteria は、通常、Phase 毎に設定されるが、今後、一定の Criteria を可能な限り複数機関で共通の数値で運用することを検討することが必要であると思われる。

検量線及び被検液の Comprehensive testing の際、一部 S.D. の高いものが見受けられたが、どの程度までであれば、採用してよいか判定が必要かもしれない。

被検液の EC<sub>50</sub> (3 回繰返し) の結果は、下記の通りであった。

表 10. 被検液の EC<sub>50</sub> (3 回繰返し) 結果一覧

Code	1st.	2nd.	3rd.	Ave.	S.D.	C.V.
H0009	0.83	0.99	0.43	0.75	0.29	38
H0010	0.27	0.26	0.21	0.25	0.032	13
H0011	3.0E-06	1.9E-06	2.6E-06	2.5E-06	5.9E-07	24
H0012	0.49	0.79	0.41	0.56	0.20	36
H0013	11	18	25	18	7.3	40
H0014	0.051	0.16	0.14	0.12	0.058	49
H0015	positive	negative	negative	-	-	-
H0016	1.4	1.3	8.1	3.6	3.95	110
H0017	positive	negative	positive	-	-	-
H0018	positive	negative	positive	-	-	-
H0019	negative	negative	negative	-	-	-
H0020	negative	positive	positive	-	-	-
H0021	550	28	18	200	300	150
H0022	25	26	27	26	0.9	3.5
H0023	negative	negative	negative	-	-	-
H0024	positive	positive	positive	-	-	-

*We need to decide up to which level of variability (C.V.) is good and acceptable. I also felt that we need protocols to confirm that the result from Graph pad prism is appropriate, and the determination of the data deleted.*

5) その他特記事項

- ・ Luminometer は、Centro LB 960, BERTHOLD TECHNOLOGIES, Germany を使用する。
- ・ 細胞は、#37~#41 継代数で行った。



4. 今後


問題となった点として、1)細胞の継代数 50 代程度までの使用期限をさらに 25 代程度までの継代数の少ない細胞使用を行うことで、突発的な活性などを極力軽減させてばらつきを無くすことが必要である。2)アッセイデータでの Criteria が個別ラボでの指標であり、測定をするたびに変動をしていくため、今後複数ラボでの共通化を図ることが必要である。


今後、上記の件を検証しながら PhaseIII 試料全 80 物質（アゴニスト試験 40 物質、アンタゴニスト試験 40 物質）の試験を繰返し 1 回の試験を行う予定である。


参考文献


- 1) LUMI-CELL ER ASSAY AGONIST PROTOCOL, National Toxicology Program (NTP) Interagency Center for the Evaluation of Alternative Toxicological Methods (NICEATM) Developed by: Xenobiotic Detection Systems, Inc. 1601 E. Geer St., Suite S Durham, NC 27704, 17 October 2008
- 2) LUMI-CELL ER ASSAY ANTAGONIST PROTOCOL, National Toxicology Program (NTP) Interagency Center for the Evaluation of Alternative Toxicological Methods (NICEATM) Developed by: Xenobiotic Detection Systems, Inc. 1601 E. Geer St., Suite S Durham, NC 27704, 17 October 2008
- 3) QA/QC の手引き(概略版)
- 4) LUMI-CELL® ER ASSAY Visual Observation Cell Viability Manual, 26 September 2007


<b>NICEATM</b> <i>National Toxicology Program Interagency Center for the Evaluation of Alternative Toxicological Methods</i>	<b>ICCVAM</b> <i>Interagency Coordinating Committee on the Validation of Alternative Methods</i>
 <h2 style="text-align: center;">Results of Phase IIb of the LUMI-CELL® ER Assay International Validation Study</h2>	
	

Overview (1)	
Slide #	
4	List of Acronyms and Abbreviations
6	The LUMI-CELL® ER Assay International Validation Study
10	The LUMI-CELL® ER Assay Validation Study - Phase I
14	The LUMI-CELL® ER Assay International Validation Study - Phase IIb
18	The LUMI-CELL® ER Assay Validation Study - Phase IIb Solubility and Range Finder Testing of Coded Agonist Test Substances
29	The LUMI-CELL® ER Assay Validation Study - Phase IIb Comprehensive Testing of Coded Agonist Substances
33	The LUMI-CELL® ER Assay Validation Study - Phase IIb Intra- and Inter-laboratory Reproducibility of Agonist Reference Standard and Controls
51	The LUMI-CELL® ER Assay Validation Study - Phase IIb Intra- and Inter-laboratory Reproducibility of Testing Results for Agonist Substances
76	The LUMI-CELL® ER Assay International Validation Study - Phase IIb ER Antagonist Solubility and Range Finder Testing
2	

Overview (2)	
Slide #	
87	The LUMI-CELL® ER Assay International Validation Study - Phase IIb ER Antagonist Comprehensive Testing
91	The LUMI-CELL® ER Assay Validation Study - Phase IIb Intra- and Inter-laboratory Reproducibility of Antagonist Reference Standard and Controls
111	The LUMI-CELL® ER Assay Validation Study - Phase IIb Intra- and Inter-laboratory Reproducibility of Testing Results for Antagonist Substances
136	Concordance of Current LUMI-CELL® ER Assay Validation Phase IIb Test Substance Results and the ICCVAM Meta-Data
145	Evaluation of LUMI-CELL® ER Phase IIb Test Failure Rates
160	Summary of Phase IIb of the LUMI-CELL® ER Assay International Validation Study
166	Progression to Phase IIb of the LUMI-CELL® ER Assay International Validation Study
3	

List of Acronyms and Abbreviations (1)	
ANOVA	Analysis of Variance
API	Apigenin
ATZ	Atrazine
BBP	Bis(2-ethylhexyl)phthalate
CASRN	Chemical Abstracts Service Registry Number
CORT	Corticosterone
CV	Coefficient of Variation
DDT	o,p-DDT
DMSO	Dimethyl sulfoxide
E2	17 $\beta$ -Estradiol
EC <sub>50</sub>	Half-maximal effective concentration
EE	Ethinyl estradiol
ER	Estrogen receptor
ECVAM	European Centre for the Validation of Alternative Methods
4	

List of Acronyms and Abbreviations (2)	
FLA	Flavone
GEN	Genistein
Hiyoshi	Hiyoshi Corporation
IC <sub>50</sub>	Concentration of substance that inhibits the reference estrogen response by 50%
ICCVAM	Interagency Coordinating Committee on the Validation of Alternative Methods
JaCVAM	Japanese Center for the Validation of Alternative Methods
NICEATM	National Toxicology Program (NTP) Interagency Center for the Evaluation of Alternative Methods
NON	p-nonylphenol
o,p-DDT	1,1,1-Trichloro-2-(p-chlorophenyl)-2-(p-chlorophenyl)ethane
Ral	Raloxifene HCl
5	

List of Acronyms and Abbreviations (3)	
RES	Resveratrol
RLU	Relative Light Units
SD	Standard Deviation
TAM	Tamoxifen
TA	Transcriptional Activation
VIN	Vinclozolin
XDS	Xenobiotic Detection Systems, Inc.
6	

## The LUMI-CELL® ER Assay International Validation Study

7

Dr. Neil Clarke, Science of Excellence

ECVAM  
NICEATM

## The International Validation Study Design

- A four phase international validation study to evaluate the reproducibility and accuracy of the LUMI-CELL® ER assay was organized by NICEATM, ECVAM, and JaCVAM.
- The study uses three laboratories, one each in the United States, Europe, and Japan.
- The study includes:
  - An evaluation of the ability of the standardized LUMI-CELL® ER assay (agonist and antagonist) protocols developed at XDS to be transferred to other laboratories.
  - An opportunity between phases of the validation study for protocol refinement.
  - Testing of 78 coded ICCVAM-recommended test substances
  - Evaluation of assay performance (comparison of results against the published literature and intra- and inter-laboratory reproducibility).

8

Dr. Neil Clarke, Science of Excellence

ECVAM  
NICEATM

## LUMICELL® ER Assay Validation Study: Study Management Team and Participating Labs

### Study Management Team:

#### ■ NICEATM

William S. Stokes, D.V.M., D.A.C.L.A.M. (NIDDK/NIH) Chair  
Raymond Tice, Ph.D. (NIDDK/NIH) Co-Chair  
Frank Deuel (S.B. Inc.) Project Coordinator  
Patricia Cayer (S.B. Inc.) Asst. Project Coordinator  
David Allen, Ph.D. (S.B. Inc.) Principal Investigator NICEATM Support Contract

#### ■ ECVAM

Susanne Bruner, Ph.D.

#### ■ JACVAM

Yoshio Kojima, Ph.D.

Hiroyuki Otsu, Ph.D.

### Participating Laboratories:

#### ■ Xenobiotic Detection Systems, Inc. (Lead Laboratory), Durham, North Carolina, U.S.

John Sorenson, Ph.D. Study Director

#### ■ ECVAM Internal Laboratory, Inra, Italy

Patrizia Pavesi, Ph.D. Study Director

Jan de Lamer

#### ■ Hiyoshi Corporation, Omi Nachiman, Japan

Hiroyuki Otsu, Ph.D. Study Director

Hiroyuki Nishida

9

Dr. Neil Clarke, Science of Excellence

ECVAM  
NICEATM

## International Validation Study Testing Phases and Timelines

STUDY PHASE	ACTIVITY	ORIGINAL TIMELINE	CURRENT TIMELINE
Phase I	Each laboratory conducts multiple testing of reference standards and controls to demonstrate proficiency with agonist and antagonist protocols. establish historical database to be used to estimate assay variability effects of test substance in Phase II, and to provide immediate or calculated reference standard and control data for an evaluation of intra- and inter-laboratory reproducibility.	Jan 07 - May 07	Nov 07 - Feb 08
Phase IIa	Four substances each from the ECVAM recommended ER inventory to be tested independently by each laboratory three times for agonist and antagonist activity.	Jun 07 - Jul 07	May 08 - Sep 08
Phase IIb	Eight substances each from the ECVAM recommended ER inventory to be tested independently by each laboratory three times for agonist and antagonist activity.	Aug 07 - Oct 07	Sep 08 - Jan 09
Phase III	Repeating all substances from ECVAM recommended ER inventory but tested once by each laboratory for agonist and antagonist activity.	Nov 07 - Dec 07	Jan 09 - Mar 09
Phase IV	Repeating all substances from ECVAM recommended ER list tested once each by the lead laboratory only for agonist and antagonist activity.	Jan 07 - Feb 07	Apr 09

10

Dr. Neil Clarke, Science of Excellence

ECVAM  
NICEATM

## The LUMI-CELL® ER Assay Validation Study - Phase I

11

Dr. Neil Clarke, Science of Excellence

ECVAM  
NICEATM

## Overview of Phase I Activities

- XDS evaluated new range finder and comprehensive test plate designs that were modified to improve testing efficiency.
- XDS, ECVAM, and Hiyoshi conducted multiple testing of agonist and antagonist protocol reference standards and controls using standardized protocols to:
  - Demonstrate proficiency with agonist and antagonist protocols
  - Demonstrate intra- and inter-laboratory reproducibility
  - Develop quality control criteria for Phase IIb testing from a historical database established from the Phase I and IIa testing of reference standards and controls

12

Dr. Neil Clarke, Science of Excellence

ECVAM  
NICEATM