

Phase I Results and Conclusions (1)

- Statistically significant differences were observed in intra- and inter-laboratory reference standard and control values.
- It was not possible to identify the causes for these differences but some of the contributing factors may be:
 - Lot-to-lot differences in cell culture media and tissue culture supplies (for intra- and inter-lab differences)
 - Differences in luminometers (for inter-lab differences)
- This underscores the importance of developing an historical control database for each individual laboratory.

13

Do Not Cite, Quote or Paraphrase



Phase I Results and Conclusions (2)

- Factors supporting reliability of the assay:
 - Assay responds robustly to E2 reference estrogen and raloxifene reference anti-estrogen.
 - Assay consistently responds to weak-acting positive controls at concentrations several orders of magnitude higher than the reference estrogen or anti-estrogen.
 - Assay plate induction (agonist) or reduction values (antagonist) were consistently greater than three-fold (only 2 of 84 plates tested had values below three-fold)
 - Phase I testing of reference standards and controls established historical databases that produced comparable test plate acceptance criteria for Phase IIb testing

14

Do Not Cite, Quote or Paraphrase



The LUMI-CELL® ER Assay Validation Study - Phase IIa

15

Do Not Cite, Quote or Paraphrase



Phase IIa Testing of Reference Standards, Controls and Coded Test Substances

- Range finder and comprehensive testing of coded test substances was conducted to:
 - Demonstrate proficiency with the agonist protocol
 - Provide reference standard, control, and coded test substance data for an evaluation of intra- and inter-laboratory reproducibility
 - Continue to build historical databases that will be used to develop acceptance criteria for tests to be conducted in Phase IIb (reference standard and control data from Phase IIb will be added to historical databases established after Phase I testing)

16

Do Not Cite, Quote or Paraphrase



Phase IIa Results and Conclusions (1)

- Four coded test substances covering a range of ER agonist and antagonist activities were each tested in at least three independent experiments for both agonist and antagonist protocols at each of the participating laboratories
 - BPA, BPB, and EE were reproducibly classified as estrogenic agonists in all of the participating laboratories
 - DDT tested as negative for estrogenic agonism at Hiyoshi (3/3), and at XDS (2/3), but was positive in all agonist tests conducted at ECVAM (3/3)
 - OBA, PROG, and TAM were reproducibly classified as estrogenic antagonists in all of the participating laboratories
 - NON was reproducibly classified as an estrogenic antagonist at ECVAM (3/3) and Hiyoshi (3/3), but was negative at XDS (3/3)

17

Do Not Cite, Quote or Paraphrase



Phase IIa Results and Conclusions (2)

- Agonist and Antagonist reference standards and controls were evaluated for intra- and inter-laboratory reproducibility
 - Reference standard and control data were reproducible within laboratories
 - Statistically significant differences were observed in intra- and inter-laboratory reference standard and control values.
 - It was not possible to identify the causes for these differences but some of the contributing factors may be:
 - Lot-to-lot differences in cell culture media and tissue culture supplies (for intra- and inter-lab differences)
 - Differences in luminometers (for inter-lab differences)
 - This underscores the importance of developing and maintaining an historical control database for each individual laboratory

18

Do Not Cite, Quote or Paraphrase



Evaluation of Phase IIa Acceptance Criteria and Failure Rates

- The failure rates of plates used for the Phase IIa comprehensive testing of coded agonist and antagonist substances were evaluated. The percentages of agonist and antagonist test plates that failed acceptance criteria across the three participating laboratories were 61% (33/54) and 38% (13/34), respectively.
 - At Hiyoshi Corporation, 11% (1/9) of agonist test plates and 0% (0/6) of antagonist test plates failed acceptance criteria
 - At XDS, Inc., 53% (8/15) of agonist plates and 43% (6/14) of antagonist test plates failed acceptance criteria
 - At the ECVAM laboratory, 80% (24/30) of agonist test plates and 50% (7/14) of antagonist test plates failed acceptance criteria

19

Do Not Copy, Quote, or Distribute

ECVAM
MOL/DA/10

Qualitative and Quantitative Evaluation of Acceptance Criteria

- To determine if changes to the acceptance criteria used for Phase IIa testing could reduce the failure rates of comprehensive test plates without compromising the ability of the assay to detect and quantify test substance agonist or antagonist activity, the qualitative (i.e., classification as an agonist or antagonist) and quantitative (i.e., agonist EC₅₀ or antagonist IC₅₀ value, if one could be calculated) outcomes for test plates that met all acceptance criteria versus those that failed to meet one or more criterion were compared.
- Results of the comparison indicated that acceptance criteria could be modified to improve failure rates without compromising the ability of the assay to detect and quantify test substance agonist or antagonist activity.

20

Do Not Copy, Quote, or Distribute

ECVAM
MOL/DA/10

Modifications Made to Test Plate Acceptance Criteria (1)

- Based on the results of the qualitative and quantitative comparisons, the SMT approved of the following changes to be used for agonist and antagonist test plate acceptance criteria for use in Phase IIb testing:
 - Agonist Test Plate Acceptance Criteria:
 - The mean plate dimethyl sulfoxide (DMSO) control relative light unit (RLU) value must be within 2.5 times the standard deviation (SD) of the historical mean RLU value for the DMSO control.
 - Plate induction, as measured by dividing the highest mean estradiol (E2) reference RLU value by the mean DMSO control RLU value, must be greater than three-fold.
 - The E2 reference standard curve should be sigmoidal in shape and have at least three values within the linear portion of the curve.
 - The mean plate methoxychlor control RLU value must be greater than three times the SD of the mean plate RLU value of the DMSO control.

21

Do Not Copy, Quote, or Distribute

ECVAM
MOL/DA/10

Modifications Made to Test Plate Acceptance Criteria (2)

- Antagonist Test Plate Acceptance Criteria:
 - The mean plate dimethyl sulfoxide (DMSO) control relative light unit (RLU) value must be within 2.5 times the standard deviation (SD) of the historical mean RLU value for the DMSO control.
 - Plate reduction, as measured by dividing the highest mean estradiol (E2) reference RLU value by the mean DMSO control RLU value, must be greater than three-fold.
 - The R_{max}E2 reference standard curve should be sigmoidal in shape and have at least three values within the linear portion of the curve.
 - The mean E2 control RLU value must be within 2.5 times the standard deviation (SD) of the historical mean RLU value for the E2 control.
 - The mean plate flutamideE2 control RLU value must be greater than three times the SD of the mean plate RLU value of the DMSO control.
- If these revised criteria had been used during Phase IIa, the overall rate of plate acceptance would have increased from 39% (21/54) to 80% (43/54) for agonist test plates, and from 62% (21/34) to 85% (29/34) for antagonist test plates.

22

Do Not Copy, Quote, or Distribute

ECVAM
MOL/DA/10

The LUMI-CELL® ER Assay Validation Study - Phase IIb

23

Do Not Copy, Quote, or Distribute

ECVAM
MOL/DA/10

Overview of Phase IIb of the Validation Study (1)

- Eight coded substances from the ER minimum list were tested for agonism and eight were tested for antagonism in each laboratory.
- The solubility of each coded substance in a 1% v/v solution of DMSO in cell culture media was determined (up to a limit concentration of 1 mg/mL).
- Following a single range finder test, at least three successful independent comprehensive tests were conducted for each coded substance.
- Test results for reference standards, controls, and coded substances were evaluated for intra- and inter-laboratory reproducibility.
- Test results for coded substances were evaluated for accuracy with agonist and antagonist activities vs. the ICCVAM meta-data

24

Do Not Copy, Quote, or Distribute

ECVAM
MOL/DA/10

Overview of Phase IIb of the Validation Study (2)

- All concentrations of coded substances that were tested in range finder or comprehensive tests were evaluated for cytotoxicity using the visual observation method developed in the LUMI-CELL® ER Protocol Validation study.

Viability Score	Description
1	Normal Cell Morphology and Cell Density
2	Altered Cell Morphology and/or Small Gaps between Cells
3	Altered Cell Morphology and/or Large Gaps between Cells
4	Few (or No) Visible Cells
P	Wells containing Precipitation are noted with "P"

25

Do Not Copy, Quote or Distribute

ECCM

NCS 12

The LUMI-CELL® ER Assay Validation Study - Phase IIb

Solubility and Range Finder Testing of Coded Agonist Substances

26

Do Not Copy, Quote or Distribute

ECCM

NCS 12

Phase IIb Agonist Coded Substances Tested

Substance Name	CABRN	Code (XDB)	Code (ECVAM)	Code (Physioli)
Alzotone (ATZ)	1913-24-9	X0015	V0006	H0015
α,β-DGT* (DGT)	786-02-6	X0011	V0008	H0010
17-β-estradiol estradiol (EE)	57-83-8	X0016	V0012	H0011
Flaxone (FLA)	525-82-6	X0013	V0008	H0012
Genistein (GEN)	446-72-0	X0012	V0007	H0014
β-nonylphenol (NON)	104-40-5	X0010	V0010	H0016
Butybenzyl sulfide (BBP)	85-68-7	X0014	V0011	H0008
Vinclozolin (VIN)	30471-44-8	X0008	V0005	H0013

*α,β-DGT = 1,1'-[Trifluoro-2-(p-chlorophenyl)-2-(p-chlorophenyl)ethane]

27

Do Not Copy, Quote or Distribute

ECCM

NCS 12

Maximum Solubility of Agonist Test Substances in Culture Media with DMSO (1%)

Test Substance	XDB		ECVAM		Physioli	
	(µg/mL)	(µM)	(µg/mL)	(µM)	(µg/mL)	(µM)
ATZ	100	464	100	464	100	464
BBP	100	320	10	32	12	32
DGT	100	262	100	262	10	26
EE	100	237	100	237	10	34
FLA	100	400	100	400	100	400
GEN	100	370	100	370	100	370
NON	100	404	10	40	100	404
VIN	100	280	10	28	100	280

28

Do Not Copy, Quote or Distribute

ECCM

NCS 12

Agonist Range Finder Testing (1)

- The purpose of range finder testing is to select starting concentrations for comprehensive testing, which is based on the following.
 - If results suggest that the test substance is negative, comprehensive testing is conducted using an 11 point 1:2 serial dilution starting at either the limit concentration or highest soluble concentration
 - If results suggest that the test substance is negative and the higher concentrations in the range finder are cytotoxic, comprehensive testing is conducted using an 11 point 1:2 serial dilution starting at the lowest cytotoxic concentration

29

Do Not Copy, Quote or Distribute

ECCM

NCS 12

Agonist Range Finder Testing (2)

- If results suggest that the test substance is positive, comprehensive testing is conducted using a concentration one log higher than the concentration giving the highest RLU value in the range finder as the starting concentration using either a 1:2 or 1:5 11-point serial dilution scheme based on:
 - An 11-point 1:2 serial dilution is used if the resulting concentration range will encompass the full range of responses based on the concentration response curve generated in the range finder test
 - If the concentration range that would be generated with the 1:2 serial dilution will not encompass the full range of responses based on the concentration response curve in the range finder test, an 11-point 1:5 serial dilution should be used instead
- If a substance exhibits a biphasic concentration response curve both curves are to be resolved using an 11-point 1:5 serial dilution with a top concentration one log higher than the concentration giving the highest RLU of the peak associated with the higher concentration in the range finder.

30

Do Not Copy, Quote or Distribute

ECCM

NCS 12

Agonist Range Finder Concentrations Tested and Cell Viability (ATZ and BBP)

Substance Name	Concentration (µg/ml)	EC50		
		ATZ	BBP	µg/ml
ATZ	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1
BBP	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1

NT = not tested

31

EC50
MCC-170

Agonist Range Finder Concentrations Tested and Cell Viability (DDT and EE)

Substance Name	Concentration (µg/ml)	EC50		
		DDT	EE	µg/ml
DDT	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1
EE	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1

NT = not tested

32

EC50
MCC-170

Agonist Range Finder Concentrations Tested and Cell Viability (FLA and GEN)

Substance Name	Concentration (µg/ml)	EC50		
		FLA	GEN	µg/ml
FLA	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1
GEN	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1

NT = not tested

33

EC50
MCC-170

Agonist Range Finder Concentrations Tested and Cell Viability (NON and VIN)


Substance Name	Concentration (µg/ml)	EC50		
		NON	VIN	µg/ml
NON	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1
VIN	1.0 x 10 ⁻⁷	1	1	1
	1.0 x 10 ⁻⁶	1	1	1
	1.0 x 10 ⁻⁵	1	1	1
	1.0 x 10 ⁻⁴	1	1	1
	1.0 x 10 ⁻³	1	1	1
	1.0 x 10 ⁻²	1	1	1

NT = not tested

34

EC50
MCC-170


XDS Agonist Range Finder Results (1)



35

EC50
MCC-170


XDS Agonist Range Finder Results (2)



36

EC50
MCC-170

ECVAM Agonist Range Finder Results (1)




37

The North Carolina State University

ECVAM
MCE/PA

ECVAM Agonist Range Finder Results (2)




38

The North Carolina State University

ECVAM
MCE/PA

Hiyoshi Agonist Range Finder Results (1)




39

The North Carolina State University

ECVAM
MCE/PA

Hiyoshi Agonist Range Finder Results (2)



40

The North Carolina State University

ECVAM
MCE/PA

The LUMI-CELL® ER Assay Validation Study - Phase IIb

Comprehensive Testing of Coded Agonist Substances

41

The North Carolina State University

ECVAM
MCE/PA

Overview of Phase IIb Comprehensive Agonist Testing

- Based on starting concentrations determined by range finder testing, at least three successful independent comprehensive agonist tests were conducted for ATZ, BBP, DDT, EE, FLA, GEN, NON, and VIN.
- Results from the successful testing of reference standard, controls, and the eight agonist substances were evaluated for intra- and inter-laboratory reproducibility.
- Test results for coded substances were evaluated for accuracy vs. agonist activities from the ICCVAM meta-data.
- Test plate failure rates were determined.

42

The North Carolina State University

ECVAM
MCE/PA

Acceptance Criteria for Phase IIb Agonist Comprehensive Testing

- Acceptance or rejection of comprehensive agonist tests conducted in Phase IIb was based on evaluation of test plate reference standards and controls. Results were evaluated against acceptance criteria derived from the historical databases established from Phase I and Phase IIb testing at each laboratory. Agonist test plate acceptance criteria used for Phase IIb are summarized as follows:
 - Plate induction, the averaged highest E2 reference standard RLU value divided by the averaged DMSO control value, must be greater than three-fold
 - DMSO control RLU values must be within 2.5 times the SD of the historical DMSO control value
 - The E2 reference standard concentration-response curve should be sigmoidal in shape and have at least three values within the linear portion of the concentration-response curve
 - Methoxychlor control RLU values must be above the line representing the DMSO mean plus three times the standard deviation from the DMSO mean

43

Do Not Use Graph or Textbox

ECVM
MCEM

The LUMI-CELL® ER Assay Validation Study - Phase IIb

Intra- and Inter-laboratory Reproducibility of Agonist Reference Standard and Controls

44

Do Not Use Graph or Textbox

ECVM
MCEM

Intralaboratory Reproducibility of Agonist Reference Standards and Controls


- Intralaboratory reproducibility of the RLU values associated with the DMSO control wells, and the fold-induction of E2 at its maximum response from test plates that met acceptance criteria was analyzed.
 - Means and standard deviations of reference standard and control values from each laboratory were evaluated and graphed
 - Linear regression analyses were conducted to assess intralaboratory reproducibility of reference standard and control results over time for each laboratory

45

Do Not Use Graph or Textbox

ECVM
MCEM

Reference Standard and Control Results from Agonist Plates Tested at XDS




46

Do Not Use Graph or Textbox

ECVM
MCEM

Reference Standard and Control Results from Agonist Plates Tested at ECVAM




47

Do Not Use Graph or Textbox

ECVM
MCEM

Reference Standard and Control Results from Agonist Plates Tested at Hiyoshi




48

Do Not Use Graph or Textbox

ECVM
MCEM

DMSO Linear Regression Analysis




	N ¹	Intercept ²	Slope	P-value (slope) ³
XDS	26	1700	22	0.47
ECVAM	26	3500	30	0.07
HIYOSHI	27	6000	12	0.42

¹Number of plates tested (combined agonist and antagonist plates)
²Intercept values are reported as unadjusted relative light units
³Statistically significant from zero at p < 0.05

40 Do Not Use, Guide to Dose-Response

ECVAM
MCC/19

Agonist E2 Fold-Induction Linear Regression Analysis



	N ¹	Intercept ²	Slope	P-value (slope) ³
XDS	13	6.7	-0.22	0.089
ECVAM	12	4.5	0.01	0.950
HIYOSHI	13	4.2	-0.01	0.987

¹Number of plates tested
²Intercept values are reported as fold-induction
³Statistically significant from zero at p < 0.05

50 Do Not Use, Guide to Dose-Response

ECVAM
MCC/19


Interlaboratory Reproducibility of Reference Standard and Controls

- Interlaboratory reproducibility of the RLU values associated with the DMSO control wells and the fold-induction of E2 at its maximum response from test plates that meet acceptance criteria was evaluated:
 - Means, standard deviations and coefficients of variation (CV) of reference standard and control values were compared
 - Variability of reference standard and control values across laboratories was evaluated by conducting an analysis of variance (ANOVA)
- If a significant p-value was obtained for the ANOVA, a Newman-Keuls post-test was used to test for significant differences in reference standard and control values between pairs of laboratories.

51 Do Not Use, Guide to Dose-Response

ECVAM
MCC/19


Interlaboratory Comparison of Reference Standard and Controls



52 Do Not Use, Guide to Dose-Response

ECVAM
MCC/19

Interlaboratory Comparison of DMSO Control



Data points represent DMSO control RLU values from agonist and antagonist plates tested in Phase III. Solid horizontal lines represent the mean DMSO control RLU value for each data set. Dashed lines indicate the mean DMSO control RLU (horizontal dashed line) value plus and minus 2.3 times the standard deviation from the mean. DMSO control RLU values cannot be less than zero.

	Plates Tested		Mean (RLU)		SD (RLU)		CV	
	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III
XDS	12	26	2261	2000	1640	1025	73%	51%
ECVAM	10	26	2929	4197	1835	1007	63%	24%
Hiyoshi	13	27	4446	6272	1789	1068	40%	17%

53 Do Not Use, Guide to Dose-Response

ECVAM
MCC/19

Interlaboratory Comparison of E2 Fold-Induction



Data points represent E2 maximum fold-induction values from plates tested in Phase III. Solid horizontal lines represent the E2 fold-induction value for each data set. Flashes are reported if the fold-induction for the maximum E2 response is less than three.

	Plates Tested		Mean (RLU)		SD (RLU)		CV	
	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III	Ph. III
XDS	7	13	4.7	7.3	3.7	2.0	43%	38
ECVAM	6	12	8.1	4.6	1.8	0.8	24%	13
Hiyoshi	8	13	4.5	4.0	0.7	0.7	15%	18

54 Do Not Use, Guide to Dose-Response

ECVAM
MCC/19

Agonist ANOVA - Interlaboratory Comparison of Reference Standard & Control Values

	p value ^{1,2}	F value ³
DMSO	<0.001	41.8
Fold-Induction	<0.001	21.8

Variability is statistically significant at p<0.05.
 MANOVA analyzed values from the three participating laboratories.
 Values in table have p-values that are less than 0.05.
¹ = ratio of between-laboratory variability to within-laboratory variability. A ratio of 1.0 indicates that the within-laboratory variability is equal and a ratio of zero indicates that all means are equal.

55

Dr. Neil Cole, Biogen, in collaboration with

ECVAM
NCEM

Newman-Keuls - Interlaboratory Comparison of Agonist Reference Standard & Controls

	DMSO ¹		Fold-Induction	
	Mean Difference	p value ^{2,3}	Mean Difference	p value ^{2,3}
XDS vs ECVAM	2228	<0.001	2.7	<0.001
XDS vs Hiyoshi	4780	<0.001	3.3	<0.001
ECVAM vs Hiyoshi	1961	<0.001	0.6	<0.05

Measured in relative light units.
 Variability is statistically significant at p<0.05.
 Values in table have p-values that are less than 0.05.

56

Dr. Neil Cole, Biogen, in collaboration with

ECVAM
NCEM

The LUMI-CELL[®] ER Assay Validation Study - Phase IIb

Intra- and Inter-laboratory Reproducibility of Testing Results for Agonist Substances

57

Dr. Neil Cole, Biogen, in collaboration with

ECVAM
NCEM

Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for ATZ

Substance Name	XDS		ECVAM		Hiyoshi	
	Concentration (ng/ml)	Cell Viability (%)	Concentration (ng/ml)	Cell Viability (%)	Concentration (ng/ml)	Cell Viability (%)
ATZ	1.00 x 10 ⁻⁷	100	1.00 x 10 ⁻⁷	100	1.00 x 10 ⁻⁷	100
	3.00 x 10 ⁻⁷	100	3.00 x 10 ⁻⁷	100	3.00 x 10 ⁻⁷	100
	1.00 x 10 ⁻⁶	100	1.00 x 10 ⁻⁶	100	1.00 x 10 ⁻⁶	100
	3.00 x 10 ⁻⁶	100	3.00 x 10 ⁻⁶	100	3.00 x 10 ⁻⁶	100
	1.00 x 10 ⁻⁵	100	1.00 x 10 ⁻⁵	100	1.00 x 10 ⁻⁵	100
	3.00 x 10 ⁻⁵	100	3.00 x 10 ⁻⁵	100	3.00 x 10 ⁻⁵	100
	1.00 x 10 ⁻⁴	100	1.00 x 10 ⁻⁴	100	1.00 x 10 ⁻⁴	100
	3.00 x 10 ⁻⁴	100	3.00 x 10 ⁻⁴	100	3.00 x 10 ⁻⁴	100
	1.00 x 10 ⁻³	100	1.00 x 10 ⁻³	100	1.00 x 10 ⁻³	100
	3.00 x 10 ⁻³	100	3.00 x 10 ⁻³	100	3.00 x 10 ⁻³	100


¹ The cell viability score indicates the score that was given for that concentration in all plates tested.

58

Dr. Neil Cole, Biogen, in collaboration with

ECVAM
NCEM

Phase IIb Agonist Results for ATZ at XDS




Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

59

Dr. Neil Cole, Biogen, in collaboration with

ECVAM
NCEM

Phase IIb Agonist Results for ATZ at ECVAM




60

Dr. Neil Cole, Biogen, in collaboration with

ECVAM
NCEM

Phase IIB Agonist Results for ATZ at Hiyoshi



B1 Do Not Copy, Share, or Distribute

Interlaboratory Comparison of ATZ Agonist Activity

	Plates Tested	Plates Testing Positive for Agonism	Plates Testing Negative for Agonism	ECVAM Meta-data
XDS	3	1	2	Negative
ECVAM	3	3	0	
Hiyoshi	3	1	2	

If mean adjusted RLU values for a given concentration(s) of test substance is greater than the mean plus three times the standard deviation of the plate (DMG) control values, the test substance is considered positive for agonist activity - any response below the threshold is considered negative for agonist activity.

B2 Do Not Copy, Share, or Distribute


Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for BBP

Reference Name	XDS		ECVAM		Hiyoshi	
	Concentration (ng/mL)	Cell Viability ¹	Concentration (ng/mL)	Cell Viability ¹	Concentration (ng/mL)	Cell Viability ¹
BBP	1.0×10^{-7}	3.22	1.0×10^{-7}	1	1.0×10^{-7}	1
	3.16×10^{-7}	1.81	3.16×10^{-7}	1	3.16×10^{-7}	1
	1.0×10^{-6}	1	1.0×10^{-6}	1	1.0×10^{-6}	1
	3.16×10^{-6}	1	3.16×10^{-6}	1	3.16×10^{-6}	1
	1.0×10^{-5}	1	1.0×10^{-5}	1	1.0×10^{-5}	1
	3.16×10^{-5}	1	3.16×10^{-5}	1	3.16×10^{-5}	1
	1.0×10^{-4}	1	1.0×10^{-4}	1	1.0×10^{-4}	1
	3.16×10^{-4}	1	3.16×10^{-4}	1	3.16×10^{-4}	1
	1.0×10^{-3}	1	1.0×10^{-3}	1	1.0×10^{-3}	1
	3.16×10^{-3}	1	3.16×10^{-3}	1	3.16×10^{-3}	1

¹The cell viability score indicates the score that was given for that concentration of all plates tested.

B3 Do Not Copy, Share, or Distribute


Phase IIB Agonist Results for BBP at XDS



Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.


B4 Do Not Copy, Share, or Distribute

Phase IIB Agonist Results for BBP at ECVAM



B5 Do Not Copy, Share, or Distribute

Phase IIB Agonist Results for BBP at Hiyoshi



B6 Do Not Copy, Share, or Distribute

Interlaboratory Comparison of BBP Agonist Activity

	Plates Tested	Plates Testing Positive for Agonist ¹	Plates Testing Negative for Agonist ¹	EC ₅₀ Meta-Data ²
XDS	3	3	0	
ECVM	3	3	0	**
Hiyoshi	3	3	0	

¹ Mean adjusted RLU values for a given concentration(s) of test substance is greater than the mean plus three times the standard deviation of the plate OMBD control values. The test substance is considered positive for agonist activity - any response below this threshold is considered negative for agonist activity

² ** indicates that the substance is classified as being moderately active in the EC₅₀ meta-data (EC₅₀ value was between 0.0001 and 0.1 μM)

67

No. Test/No. Plates or Dishes

ECVM
MCE/IN

Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for DDT

Substance Name	XDS		ECVM		Hiyoshi	
	Concentration (μg/mL)	Cell Viability ¹	Concentration (μg/mL)	Cell Viability ¹	Concentration (μg/mL)	Cell Viability ¹
DDT	1.00 x 10 ⁻¹	3	1.00 x 10 ⁻¹	3	1.00 x 10 ⁻¹	3
	0.30 x 10 ⁻¹	3	0.30 x 10 ⁻¹	3	0.30 x 10 ⁻¹	3
	0.03 x 10 ⁻¹	3	0.03 x 10 ⁻¹	3	0.03 x 10 ⁻¹	3
	1.00 x 10 ⁻²	3	0.03 x 10 ⁻¹	3	0.03 x 10 ⁻¹	3
	0.30 x 10 ⁻²	3	1.00 x 10 ⁻²	3	1.00 x 10 ⁻²	3
	0.10 x 10 ⁻²	3	0.03 x 10 ⁻¹	3	0.03 x 10 ⁻¹	3
	1.00 x 10 ⁻³	3	0.03 x 10 ⁻¹	3	0.03 x 10 ⁻¹	3
	0.30 x 10 ⁻³	3	1.00 x 10 ⁻³	3	1.00 x 10 ⁻³	3
	0.10 x 10 ⁻³	3	0.03 x 10 ⁻¹	3	0.03 x 10 ⁻¹	3
	1.00 x 10 ⁻⁴	3	0.10 x 10 ⁻³	3	0.10 x 10 ⁻³	3
0.30 x 10 ⁻⁴	3	1.00 x 10 ⁻⁴	3	1.00 x 10 ⁻⁴	3	


¹ Multiple scores indicate the cell viability scores for concentrations tested in each plate when scores differed from plate to plate. A single cell viability score indicates the score that was given for that concentration on all plates tested.

68

No. Test/No. Plates or Dishes

ECVM
MCE/IN

Phase IIb Agonist Results for DDT at XDS




Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

69

No. Test/No. Plates or Dishes

ECVM
MCE/IN

Phase IIb Agonist Results for DDT at ECVM




70

No. Test/No. Plates or Dishes

ECVM
MCE/IN

Phase IIb Agonist Results for DDT at Hiyoshi




71

No. Test/No. Plates or Dishes

ECVM
MCE/IN

Interlaboratory Comparison of DDT EC₅₀ Values



Data points represent DDT EC₅₀ values from plates tested in Phase IIb.

Plates Tested	Mean (μg/mL)	SD (μg/mL)	CV (%)	EC ₅₀ Meta-Data ¹
XDS	3.2 x 10 ⁻⁴	6.8 x 10 ⁻⁵	30	
ECVM	3	15.0 x 10 ⁻⁵	24.7 x 10 ⁻¹	**
Hiyoshi	3	24.7 x 10 ⁻⁵	32.5 x 10 ⁻¹	

¹ ** indicates that the substance is classified as being weakly active in the EC₅₀ meta-data (EC₅₀ value was > 0.1 μM)

72

No. Test/No. Plates or Dishes

ECVM
MCE/IN

Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for EE

Substance Name	XDS		ECVAM		Hiyoshi	
	Concentration (µg/ml)	Cell Viability (%)	Concentration (µg/ml)	Cell Viability (%)	Concentration (µg/ml)	Cell Viability (%)
EE	1.00 x 10 ⁻¹	1	1.00 x 10 ⁻¹	1	1.00 x 10 ⁻¹	1
	3.16 x 10 ⁻¹	1	1.00 x 10 ⁻¹	1	3.16 x 10 ⁻¹	1
	1.00 x 10 ⁰	1	3.16 x 10 ⁻¹	1	1.00 x 10 ⁰	1
	3.16 x 10 ⁰	1	1.00 x 10 ⁰	1	3.16 x 10 ⁰	1
	1.00 x 10 ¹	1	3.16 x 10 ⁰	1	1.00 x 10 ¹	1
	3.16 x 10 ¹	1	1.00 x 10 ¹	1	3.16 x 10 ¹	1
	1.00 x 10 ²	1	3.16 x 10 ¹	1	1.00 x 10 ²	1
	3.16 x 10 ²	1	1.00 x 10 ²	1	3.16 x 10 ²	1
	1.00 x 10 ³	1	3.16 x 10 ²	1	1.00 x 10 ³	1
	3.16 x 10 ³	1	1.00 x 10 ³	1	3.16 x 10 ³	1
	1.00 x 10 ⁴	1	3.16 x 10 ³	1	1.00 x 10 ⁴	1
	3.16 x 10 ⁴	1	1.00 x 10 ⁴	1	3.16 x 10 ⁴	1


*The cell viability score indicates the score that was given for that concentration on all plates tested.

73

De Novo Drug, Status on Demand

ECVAM
MCC/IN

Phase IIb Agonist Results for EE at XDS




*Cell concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

74

De Novo Drug, Status on Demand

ECVAM
MCC/IN

Phase IIb Agonist Results for EE at ECVAM




75

De Novo Drug, Status on Demand

ECVAM
MCC/IN

Phase IIb Agonist Results for EE at Hiyoshi




76

De Novo Drug, Status on Demand

ECVAM
MCC/IN

Interlaboratory Comparison of EE E2 EC₅₀ Values



Data points represent EE EC₅₀ values from plates tested in Phase II.

	Plates Tested	Mean (µg/mL)	SD (µg/mL)	CV (%)	ECVAM Meta-data*
XDS	3	2.1 x 10 ⁻³	7.8 x 10 ⁻⁴	36	
ECVAM	3	1.7 x 10 ⁻³	6.3 x 10 ⁻⁴	35	***
Hiyoshi	3	2.5 x 10 ⁻³	5.8 x 10 ⁻⁴	24	

**** indicates that the substance is classified as being strongly active in the ECVAM meta-data (EC₅₀ value was < 0.0001 µM).

77

De Novo Drug, Status on Demand

ECVAM
MCC/IN

Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for FLA

Substance Name	XDS		ECVAM		Hiyoshi	
	Concentration (µg/ml)	Cell Viability (%)	Concentration (µg/ml)	Cell Viability (%)	Concentration (µg/ml)	Cell Viability (%)
FLA	1.00 x 10 ⁻¹	1	1.00 x 10 ⁻¹	1	1.00 x 10 ⁻¹	1
	3.16 x 10 ⁻¹	2,2	3.16 x 10 ⁻¹	1	2.80 x 10 ⁻¹	1
	1.00 x 10 ⁰	1	1.00 x 10 ⁰	1	1.00 x 10 ⁰	1
	3.16 x 10 ⁰	1	3.16 x 10 ⁰	1	3.16 x 10 ⁰	1
	1.00 x 10 ¹	1	1.00 x 10 ¹	1	1.00 x 10 ¹	1
	3.16 x 10 ¹	1	3.16 x 10 ¹	1	3.16 x 10 ¹	1
	1.00 x 10 ²	1	1.00 x 10 ²	1	1.00 x 10 ²	1
	3.16 x 10 ²	1	3.16 x 10 ²	1	3.16 x 10 ²	1
	1.00 x 10 ³	1	1.00 x 10 ³	1	1.00 x 10 ³	1
	3.16 x 10 ³	1	3.16 x 10 ³	1	3.16 x 10 ³	1
	1.00 x 10 ⁴	1	1.00 x 10 ⁴	1	1.00 x 10 ⁴	1
	3.16 x 10 ⁴	1	3.16 x 10 ⁴	1	3.16 x 10 ⁴	1


*The cell viability score indicates the score that was given for that concentration on all plates tested.

78

De Novo Drug, Status on Demand


ECVAM
MCC/IN

Phase IIb Agonist Results for FLA at XDS




Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

Phase IIb Agonist Results for FLA at ECVAM



Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

Phase IIb Agonist Results for FLA at Hyoshi



Interlaboratory Comparison of FLA Agonist Activity

	Phase Tested	Phase Testing Positive for Agonist ¹	Phase Testing Negative for Agonist ¹	ICCVAM Meta-data ²
XDS	3	3	0	+
ECVAM	3	3	0	
Hyoshi	3	3	0	

¹ Mean adjusted RLU values for a given concentration(s) of test substance is greater than the mean plus three times the standard deviation of the plate DMSO control values, the test substance is considered positive for agonist activity; any response below this threshold is considered negative for agonist activity.


² + Indicates that the substance is classified as being weakly active in the ICCVAM meta-data (EC₅₀ value was > 0.1 µM).

Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for GEN

Substance Name	XDS		ECVAM		Hyoshi	
	Concentration (µg/ml)	Cell Viability ¹	Concentration (µg/ml)	Cell Viability ¹	Concentration (µg/ml)	Cell Viability ¹
GEN	1.00 x 10 ⁻¹	1.0	1.00 x 10 ⁻¹	1	1.00 x 10 ⁻¹	1
	0.00 x 10 ⁻¹	1.0	0.00 x 10 ⁻¹	1	0.00 x 10 ⁻¹	1
	0.00 x 10 ⁻²	1	0.00 x 10 ⁻²	1	0.00 x 10 ⁻²	1
	0.00 x 10 ⁻³	1	0.00 x 10 ⁻³	1	0.00 x 10 ⁻³	1
	0.00 x 10 ⁻⁴	1	0.00 x 10 ⁻⁴	1	0.00 x 10 ⁻⁴	1
	0.00 x 10 ⁻⁵	1	0.00 x 10 ⁻⁵	1	0.00 x 10 ⁻⁵	1
	0.00 x 10 ⁻⁶	1	0.00 x 10 ⁻⁶	1	0.00 x 10 ⁻⁶	1
	0.00 x 10 ⁻⁷	1	0.00 x 10 ⁻⁷	1	0.00 x 10 ⁻⁷	1
	0.00 x 10 ⁻⁸	1	0.00 x 10 ⁻⁸	1	0.00 x 10 ⁻⁸	1
	0.00 x 10 ⁻⁹	1	0.00 x 10 ⁻⁹	1	0.00 x 10 ⁻⁹	1
0.00 x 10 ⁻¹⁰	1	0.00 x 10 ⁻¹⁰	1	0.00 x 10 ⁻¹⁰	1	


¹ The cell viability score indicates the score that was given for that concentration in all plates tested.

Phase IIb Agonist Results for GEN at XDS



Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

Phase IIb Agonist Results for GEN at ECVAM




85

Do Not Use, Quote or Distribute

ECVAM
M03.17c

Phase IIb Agonist Results for GEN at Hiyoshi



86

Do Not Use, Quote or Distribute

ECVAM
M03.17c

Interlaboratory Comparison of GEN EC₅₀ Values



Data points represent GEN EC₅₀ values from plates tested in Phase IIb

Plate	Mean (µg/mL)	SD (µg/mL)	CV (%)	ECVAM Mean-SD*
XDS	3.7×10^{-7}	1.6×10^{-7}	20	
ECVAM	8.1×10^{-7}	3.8×10^{-7}	11	
Hiyoshi	1.6×10^{-6}	47.5×10^{-7}	40	

* Indicates that the substance is classified as being weakly active in the ECVAM meta-data (EC₅₀ value was > 0.1 µM)

87

Do Not Use, Quote or Distribute

ECVAM
M03.17c

Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for NON

Substance Name	XDS		ECVAM		Hiyoshi	
	Concentration (µg/mL)	Cell Viability (%)	Concentration (µg/mL)	Cell Viability (%)	Concentration (µg/mL)	Cell Viability (%)
NON	1.00×10^{-7}	9	1.00×10^{-7}	20, 20	1.00×10^{-7}	6
	2.00×10^{-7}	2, 3	2.00×10^{-7}	1	2.00×10^{-7}	6
	3.00×10^{-7}	1	3.00×10^{-7}	1	3.00×10^{-7}	4
	4.00×10^{-7}	1	4.00×10^{-7}	1	4.00×10^{-7}	2
	5.00×10^{-7}	1	5.00×10^{-7}	1	5.00×10^{-7}	1
	6.00×10^{-7}	1	6.00×10^{-7}	1	6.00×10^{-7}	1
	7.00×10^{-7}	1	7.00×10^{-7}	1	7.00×10^{-7}	1
	8.00×10^{-7}	1	8.00×10^{-7}	1	8.00×10^{-7}	1
	9.00×10^{-7}	1	9.00×10^{-7}	1	9.00×10^{-7}	1
	1.00×10^{-6}	1	1.00×10^{-6}	1	1.00×10^{-6}	1


Multiple scores indicate the cell viability score for concentrations tested in each plate when scores differed from plate to plate. A single cell viability score indicates the score that was greater for that concentration on all plates tested.

88

Do Not Use, Quote or Distribute

ECVAM
M03.17c

Phase IIb Agonist Results for NON at XDS




Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

89

Do Not Use, Quote or Distribute

ECVAM
M03.17c

Phase IIb Agonist Results for NON at ECVAM




Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

90

Do Not Use, Quote or Distribute

ECVAM
M03.17c

Phase IIb Agonist Results for NON at Hiyoshi



Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

91

ECVAM
MCC-19

Interlaboratory Comparison of NON Agonist Activity

	Plates Tested	Plates Testing Positive for Agonist ^a	Plates Testing Negative for Agonist ^a	ICCVAM Mean-Std ^b
XDS	3	3	0	
ECVAM	3	3	0	**
Hiyoshi	3	3	0	

^aIf mean adjusted RLJ values for a given concentration(s) of test substance is greater than the mean plus three times the standard deviation of the plate (MCC) control values, the test substance is considered positive for agonist activity; any response below this threshold is considered negative for agonist activity.

^b** Indicates that the substance is classified as being moderately active in the ICCVAM meta-data (IC₅₀ value was between 0.001 and 0.1 μM).

92

ECVAM
MCC-19

Agonist Comprehensive Testing - Concentrations Tested & Cell Viability for VIN


Substance Name	XDS		ECVAM		Hiyoshi	
	Concentration (ng/mL)	Cell Viability ^a	Concentration (ng/mL)	Cell Viability ^a	Concentration (ng/mL)	Cell Viability ^a
VIN	1.00 x 10 ⁻⁷	3.2	1.00 x 10 ⁻⁷	5	1.00 x 10 ⁻⁷	5
	2.00 x 10 ⁻⁷	3.2	4.00 x 10 ⁻⁷	5	5.00 x 10 ⁻⁷	5
	4.00 x 10 ⁻⁷	5	2.00 x 10 ⁻⁶	5	2.00 x 10 ⁻⁶	5
	8.00 x 10 ⁻⁷	5	1.00 x 10 ⁻⁵	5	1.00 x 10 ⁻⁵	5
	1.60 x 10 ⁻⁶	5	4.00 x 10 ⁻⁶	5	4.00 x 10 ⁻⁶	5
	3.20 x 10 ⁻⁶	5	1.00 x 10 ⁻⁵	5	1.00 x 10 ⁻⁵	5
	6.40 x 10 ⁻⁶	5	1.00 x 10 ⁻⁵	5	1.00 x 10 ⁻⁵	5
	1.28 x 10 ⁻⁵	5	2.00 x 10 ⁻⁵	5	2.00 x 10 ⁻⁵	5
	2.56 x 10 ⁻⁵	5	4.00 x 10 ⁻⁵	5	4.00 x 10 ⁻⁵	5
	5.12 x 10 ⁻⁵	5	8.00 x 10 ⁻⁵	5	8.00 x 10 ⁻⁵	5
1.02 x 10 ⁻⁴	5	1.60 x 10 ⁻⁴	5	3.20 x 10 ⁻⁴	5	

^aThe cell viability score indicates the score that was given for the concentration or all plates tested.

93

ECVAM
MCC-19

Phase IIb Agonist Results for VIN at XDS




Test concentrations that were cytotoxic (cell viability scores of 2 or greater) were excluded from the graph because they are not used in the determination of agonist activity.

94

ECVAM
MCC-19


Phase IIb Agonist Results for VIN at ECVAM



95

ECVAM
MCC-19

Phase IIb Agonist Results for VIN at Hiyoshi



96

ECVAM
MCC-19

Interlaboratory Comparison of VIN Agonist Activity

	Plates Tested	Plates Testing Positive for Agonism	Plates Testing Negative for Agonism	EC ₅₀ M Meta-data
KDS	3	0	3	Negative
EC ₅₀ M	3	3	0	
Hydrol	4	1	3	

If mean adjusted RLU values for a given concentration(s) of test substance is greater than the mean plus three times the standard deviation of the plate DMSO control values, the test substance is considered positive for agonist activity - any response below this threshold is considered negative for agonist activity.

The LUMI-CELL® ER Assay Validation Study - Phase IIb

Solubility and Range Finder Testing of Coded Antagonist Substances

98

By Herbig, Bock, et al. (2016)

ECVAM
MCSA No.

Phase IIb Antagonist Coded Substances Tested

Substance Name	CASRN	Code (DOB)	Code (ECVAM)	Code (Hyochil)
Aggrenin (API)	520-36-5	X024	V0089	0020
Atazone (ATZ)	1912-24-9	X023	V0067	H024
Corticosterone (CORT)	50-22-8	X021	V0065	H021
o,p'-DDT (DDT)	786-52-6	X018	V0064	H018
Flavone (FLA)	525-52-8	X020	V0090	H019
Genistein (GEN)	446-72-0	X019	V0063	H023
Butylbenzyl p-Hydroxy (BBP)	85-68-7	X022	V0086	H017
Resveratrol (RES)	501-36-0	X017	V0068	H022

¹o,p'-DDT = 1,1'-Dichloro-2,2'-bis(4-chlorophenyl)-2,2'-bis(4-chlorophenyl)ethane

99

By Herbig, Bock, et al. (2016)

ECVAM
MCSA No.

Maximum Solubility of Antagonist Test Substances in Culture Media with DMSO (1%)

Test Substance	DOB		ECVAM		Hyochil	
	(µg/mL)	(µM)	(µg/mL)	(µM)	(µg/mL)	(µM)
API	100	370	10	37	10	37
ATZ	100	464	100	464	100	464
BBP	100	520	10	32	10	32
CORT	1000	2888	1000	2888	100	280
DDT	100	282	10	28	10	28
FLA	100	400	100	400	10	40
GEN	100	370	100	370	10	37
RES	100	438	100	438	100	438

100

By Herbig, Bock, et al. (2016)

ECVAM
MCSA No.

Antagonist Range Finder Testing (1)

- The purpose of range finder testing is to select starting concentrations for comprehensive testing, which is based on the following:
 - If results suggest that the test substance is negative, comprehensive testing is conducted using an 11 point 1:2 serial dilution with the limit dose as the starting concentration
 - If results suggest that the test substance is negative and the higher concentrations in the range finder are cytotoxic, comprehensive testing is conducted using an 11 point 1:2 serial dilution with the lowest cytotoxic concentration as the starting concentration

101

By Herbig, Bock, et al. (2016)

ECVAM
MCSA No.

Antagonist Range Finder Testing (2)

- If results suggest that the test substance is positive, comprehensive testing is conducted using a top concentration one log higher than the concentration giving the highest RLU value in the range finder as the starting concentration using either a 1:2 or 1:5 11-point serial dilution scheme based on:
 - An 11-point 1:2 serial dilution is used if the resulting concentration range will encompass the full range of responses based on the concentration response curve generated in the range finder test.
 - If the concentration range that would be generated with the 1:2 serial dilution will not encompass the full range of responses based on the concentration response curve in the range finder test, an 11-point 1:5 serial dilution should be used instead.
- If a substance exhibits a biphasic concentration response curve both curves are to be resolved using an 11-point 1:5 serial dilution with a top concentration one log higher than the concentration giving the highest RLU of the peak associated with the higher concentration in the range finder.

102

By Herbig, Bock, et al. (2016)

ECVAM
MCSA No.

Antagonist Range Finder Concentrations Tested and Cell Viability (API and ATZ)

Substance Name	Concentration (µg/mL)	DOB		ECVAM		Hyochil	
		Cell Viability	Cell Viability	Cell Viability	Cell Viability		
API	1.0 µg/mL	+	+	+	+	+	+
	0.1 µg/mL	+	+	+	+	+	+
	0.01 µg/mL	+	+	+	+	+	+
	0.001 µg/mL	+	+	+	+	+	+
	0.0001 µg/mL	+	+	+	+	+	+
	0.00001 µg/mL	+	+	+	+	+	+
ATZ	1.0 µg/mL	+	+	+	+	+	+
	0.1 µg/mL	+	+	+	+	+	+
	0.01 µg/mL	+	+	+	+	+	+
	0.001 µg/mL	+	+	+	+	+	+
	0.0001 µg/mL	+	+	+	+	+	+
	0.00001 µg/mL	+	+	+	+	+	+

NT = not tested

103

By Herbig, Bock, et al. (2016)

ECVAM
MCSA No.

Antagonist Range Finder Concentrations Tested and Cell Viability (BBP and CORT)

Substance Name	Concentration (µg/ml)	BBP	ECVAM	Hypack
		Cell Viability	Cell Viability	Cell Viability
BBP	1.0 x 10 ⁻¹	0	NT	NT
	1.0 x 10 ⁻²	0	0	0
	1.0 x 10 ⁻³	0	0	0
	1.0 x 10 ⁻⁴	0	0	0
	1.0 x 10 ⁻⁵	0	0	0
	1.0 x 10 ⁻⁶	0	0	0
	1.0 x 10 ⁻⁷	0	0	0
CORT	1.0 x 10 ⁻¹	0	0	0
	1.0 x 10 ⁻²	0	0	0
	1.0 x 10 ⁻³	0	0	0
	1.0 x 10 ⁻⁴	0	0	0
	1.0 x 10 ⁻⁵	0	0	0
	1.0 x 10 ⁻⁶	0	0	0
	1.0 x 10 ⁻⁷	0	0	0

NT = Not tested

104

Do Not Use, Grade or Distribution

ECVAM

MSL/W

Antagonist Range Finder Concentrations Tested and Cell Viability (DDT and FLA)

Substance Name	Concentration (µg/ml)	DDT	ECVAM	Hypack
		Cell Viability	Cell Viability	Cell Viability
DDT	1.0 x 10 ⁻¹	0	NT	NT
	1.0 x 10 ⁻²	0	0	0
	1.0 x 10 ⁻³	0	0	0
	1.0 x 10 ⁻⁴	0	0	0
	1.0 x 10 ⁻⁵	0	0	0
	1.0 x 10 ⁻⁶	0	0	0
	1.0 x 10 ⁻⁷	0	0	0
FLA	1.0 x 10 ⁻¹	0	0	0
	1.0 x 10 ⁻²	0	0	0
	1.0 x 10 ⁻³	0	0	0
	1.0 x 10 ⁻⁴	0	0	0
	1.0 x 10 ⁻⁵	0	0	0
	1.0 x 10 ⁻⁶	0	0	0
	1.0 x 10 ⁻⁷	0	0	0

NT = Not tested

105

Do Not Use, Grade or Distribution

ECVAM

MSL/W

Antagonist Range Finder Concentrations Tested and Cell Viability (GEN and RES)

Substance Name	Concentration (µg/ml)	GEN	ECVAM	Hypack
		Cell Viability	Cell Viability	Cell Viability
GEN	1.0 x 10 ⁻¹	0	0	NT
	1.0 x 10 ⁻²	0	0	0
	1.0 x 10 ⁻³	0	0	0
	1.0 x 10 ⁻⁴	0	0	0
	1.0 x 10 ⁻⁵	0	0	0
	1.0 x 10 ⁻⁶	0	0	0
	1.0 x 10 ⁻⁷	0	0	0
RES	1.0 x 10 ⁻¹	0	0	0
	1.0 x 10 ⁻²	0	0	0
	1.0 x 10 ⁻³	0	0	0
	1.0 x 10 ⁻⁴	0	0	0
	1.0 x 10 ⁻⁵	0	0	0
	1.0 x 10 ⁻⁶	0	0	0
	1.0 x 10 ⁻⁷	0	0	0

NT = Not tested


106

Do Not Use, Grade or Distribution

ECVAM

MSL/W

XDS Antagonist Range Finder Results (1)




107

Do Not Use, Grade or Distribution

ECVAM

MSL/W

XDS Antagonist Range Finder Results (2)




108

Do Not Use, Grade or Distribution

ECVAM

MSL/W

ECVAM Antagonist Range Finder Results (1)






109

Do Not Use, Grade or Distribution

ECVAM

MSL/W



The LUMI-CELL® ER Assay Validation Study - Phase IIb

Antagonist Comprehensive Testing

113 Do Not Copy, Quote, or Distribute

Overview of Phase IIb Comprehensive Antagonist Testing

- Based on starting concentrations determined by range finder testing, at least three successful independent comprehensive agonist tests were conducted for API, ATZ, BBP, CORT, DDT, FLA, GEN, and RES.
- Results from the successful testing of reference standard, controls, and the eight antagonist substances were evaluated for intra- and inter-laboratory reproducibility.
- Test results for coded substances were evaluated for concordance with antagonist activities from the ICCVAM meta-data
- Test plate failure rates were determined

114 Do Not Copy, Quote, or Distribute

Acceptance Criteria for Phase IIb Antagonist Testing

- Acceptance or rejection of antagonist tests conducted in Phase IIb was based on evaluation of test plate reference standard and control results. Results were compared to acceptance criteria derived from the historical databases established from Phase I and IIa testing at each laboratory. Antagonist test plate acceptance criteria used in Phase IIb are summarized as follows:
 - Plate reduction, the averaged highest Rai/E2 reference standard RLU value divided by the averaged lowest Rai/E2 reference standard value (comprehensive testing only) must be greater than three-fold
 - DMSO control RLU values must be within 2.5 times the of the historical DMSO control value range finder and comprehensive testing)
 - E2 control RLU values must be within 2.5 times of the historical E2 control value (comprehensive testing only)
 - The Rai/E2 reference standard concentration-response curve should be sigmoidal in shape and have at least three values within the linear portion of the concentration-response curve
 - Flavone/E2 control RLU values must less than the E2 control mean minus three times the standard deviation from the E2 control mean.

115 Do Not Copy, Quote, or Distribute

The LUMI-CELL® ER Assay Validation Study - Phase IIb

Intra- and Inter-laboratory Reproducibility of Antagonist Reference Standard and Controls

116

ECVAM
NCE/DA/79


Intralaboratory Reproducibility of Antagonist Reference Standards and Controls

- Intralaboratory reproducibility of the RLU values associated with the DMSO control wells, the fold-reduction of Ral/E2 at its maximum response, and the adjusted and normalized RLU values associated with the E2 control wells were statistically analyzed.
 - Means and standard deviations of reference standard and control values from each laboratory were evaluated and graphed
 - Linear regression analyses were conducted to assess intralaboratory reproducibility of reference standard and control results over time for each laboratory
- If a significant p-value was obtained for the ANOVA, a Newman-Keuls post-test was used to test for significant differences in reference standard and control values between pairs of laboratories.

117

ECVAM
NCE/DA/79


Test Plate Results for Reference Standard from Antagonist Testing at XDS



118

ECVAM
NCE/DA/79


Test Plate Results for Reference Standard from Antagonist Testing at ECVAM



119

ECVAM
NCE/DA/79


Test Plate Results for Reference Standard from Antagonist Testing Hiyoshi



120

ECVAM
NCE/DA/79

DMSO Linear Regression Analysis




	N ^a	Intercept ^b	Slope	P-value (slope) ^c
XDS	26	1760	.22	0.47
ECVAM	24	3600	.30	0.07
HIYOSH	28	8000	.12	0.42

^aNumber of plates tested (combined against and antagonist plates)
^bIntercept values are reported as unadjusted relative light units
^cStatistically significant from zero at p < 0.05

121

ECVAM
NCE/DA/79

Ral/E2 Fold-Reduction Linear Regression Analysis




	N ^a	Intercept ^b	Slope	P-value (slope) ^c
XDS	12	10.6	0.23	0.848
ECVAM	12	6.8	-0.01	0.684
HIYOSHI	14	12.0	-0.04	0.186

^aNumber of plates tested
^bIntercept values are reported as fold-reduction
^cStatistically significant from zero at p < 0.05

122

ECVAM
M03174

E2 Control Linear Regression Analysis




	N ^a	Intercept ^b	Slope	P-value (slope) ^c
XDS	12	7944	116	0.296
ECVAM	12	10910	-58	0.031
HIYOSHI	14	5600	26	0.166

^aNumber of plates tested
^bIntercept values are reported as relative light units
^cStatistically significant from zero at p < 0.05
^dValues in bold have p values that are less than 0.05

123

ECVAM
M03174


Comparison of Antagonist Reference Standard and Controls



124

ECVAM
M03174

Interlaboratory Comparison of DMSO Control




Data points represent DMSO control RLU values from agonist and antagonist plates tested in Phase II. Solid horizontal lines represent the mean DMSO control RLU value for each data set. Dashed lines indicate the mean DMSO control RLU historical reference value plus and minus 2.5 times the standard deviation from the mean. DMSO control RLU values control to less than 20%.

	Plates Tested		Mean (RLU)		SD (RLU)		CV	
	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib
XDS	12	25	2261	2000	1640	1025	73%	51%
ECVAM	12	24	2929	4187	1835	1007	63%	24%
Hiyoshi	13	27	4448	8272	1769	1099	40%	13%

125

ECVAM
M03174

Interlaboratory Comparison of Ral/E2 Fold-Reduction




Data points represent Ral/E2 fold-reduction values from plates tested in Phase II. Solid horizontal lines represent the Ral/E2 fold-reduction value for each data set. Plates are reported if the fold-reduction for the maximum Ral/E2 response is less than 20%.

	Plates Tested		Mean (RLU)		SD (RLU)		CV	
	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib
XDS	6	12	11.1	11.4	2.68	2.44	24%	21%
ECVAM	7	12	12.1	8.57	1.67	0.84	14%	10%
Hiyoshi	6	14	11.4	10.9	3.22	1.63	28%	15%

126

ECVAM
M03174

Interlaboratory Comparison of E2 Control



Data points represent E2 control values in adjusted RLU from plates tested in Phase II. Solid horizontal lines represent the mean E2 control value for each data set. Dashed lines indicate the mean E2 control historical reference value plus and minus 2.5 times the standard deviation from the mean.

	Plates Tested		Mean (RLU)		SD (RLU)		CV	
	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib	Ph. Ia	Ph. Ib
XDS	6	12	8264	8259	793	711	9%	9%
ECVAM	7	12	8891	9175	554	725	6%	8%
Hiyoshi	6	14	6728	8270	347	478	5%	6%

127

ECVAM
M03174