presentation (WG9). 25<sup>th</sup> Meeting of ISO/TC 194 (2014.4, Mishima, Japan).

H. 知的財産権の出願・登録状況 特になし

## 参考文献

- 1) FDA: Safety Assessment of Di (2-ethylhexyl) phthalate (DEHP), Released from PVC Medical Devices. 2002.
- 2) ISO 3826-1 (2013): Plastics collapsible containers for human blood and blood components.
- 3) JIS T 3217 (2011):血液成分分離バッグ
- 4) ASTM F 756-08 (2008): Standard Practice for Assessment of Hemolytic Properties of Materials.
- 5)日比谷信 他:体外循環用血液回路からの可塑剤溶出の 基礎的検討— DEHP と TOTM の溶出の比較—. 体外循環 技術. Vol. 33, No. 12, p16-19, 2006.
- 6) Miyamoto M et al.: Effects of autoclave sterilization on the physical properties of storage bags and granulocyte function. Vox Sang. Vol. 54, p74-7, 1988.
- 7) 橋本周郎 他:ポリ塩化ビニル樹脂用可塑剤 TOTM の安全性 (DEHP との比較). 第 25 回日本バイオマテリアル学会大会予稿集, p375, 2003.
- 8) 伊藤好文 他:新規ポリ塩化ビニル樹脂用可塑剤 TOTM の簡易生殖発生毒性評価.第 25 回日本バイオマテリアル学会大会予稿集, p376, 2003.




図1. 体外循環用 TOTM-PVC 製チューブからの TOTM 溶出量 (内表面積あたり)

表1. TOTM 溶出量測定サンプルー体外循環用血液回路チューブ

No.	材質、組成、製造日、滅菌条件等
1	塩化ビニル樹脂 ( TOTM 78 パーツ )、φ 6 mm、EOG 滅菌品
2	塩化ビニル樹脂(TOTM 78 パーツ)、φ12 mm、EOG 滅菌品
3	塩化ビニル樹脂 ( TOTM 85 パーツ )、φ 6 mm、EOG 滅菌品
4	塩化ビニル樹脂(TOTM 85 パーツ)、φ12 mm、EOG 滅菌品

表2. MS分析条件

プローブ	APCI
ネブライズガス流量	2.5 L/min
APCI インターフェイス温度	400 °C
CDI 温度	250 ℃
ヒートブロック温度	200 °C
検出器電圧	1.5 kV
CDL 電圧	5.0 V

表3. 血液適合性評価サンプル(体外循環用血液回路チューブ)

No.	材質、組成、製造日、滅菌条件等
1	塩化ビニル樹脂 (TOTM 68.5 パーツ)、φ3 mm、γ 滅菌品
2	塩化ビニル樹脂 ( TOTM 68.5 パーツ)、φ3 mm、EOG 滅菌品
3	塩化ビニル樹脂 (TOTM 68.5 パーツ)、φ3 mm、電子放射線滅菌品
4	塩化ビニル樹脂 ( TOTM 85.0 パーツ)、φ3 mm、EOG 滅菌品

表4. TOTM-PVC 製チューブの溶血毒性試験結果

S4 - 31 - 43	ASTM Hemolytic Index (%)		
Sterilization —	Extraction Method	Direct contact method	
Gamma-ray	0.8	2.4	
EOG	0.0	0.0	
Electron Beam	0.4	0.9	
EOG	0.0	0.2	
	EOG Electron Beam	SterilizationExtraction MethodGamma-ray0.8EOG0.0Electron Beam0.4	

# Ⅱ. 学会等発表実績

成果報告一覧

## 学 会 等 発 表 実 績

委託業務題目「医薬品・医療機器の実用化促進のための評価技術手法の戦略的開発」 機関名 国立医薬品食品衛生研究所 医療機器部

### 1. 学会等における口頭・ポスター発表

発表した成果(発表題目、口 頭・ポスター発表の別)	発表者氏名	発表した場所 (学会等名)	発表した時期	国内・外の別
新規血液バッグ用代替可塑剤 DOTHのラット亜慢性毒性試験・ ポスター発表	<b>蓜福山野小熊藤井森市前高河伊柚鄭川新吉島井﨑村園田澤上川村田橋上間場 州見田由千佳祐 秀彩上朋亮 美強間俊雄久伸田工世计知文乃薫美平潤和志郎康一美吾緑二恵世介知文乃薫美平潤和志郎</b>	第36回 日本バイオマテリアル学会大会 (船堀)	2014. 11	国内
新規血液バッグ素材DOTH/DINCH 配合PVCシートの性能評価・ポスター発表		第36回 日本バイオマテリアル学会大会 (船堀)	2014. 11	国内
簡易溶血性試験法の性能評価と 公定法との比較検証・ポスター 発表	野福柚新坂谷杉竹新蓜村井場藤口川山内見島休千俊智圭隆知美再日中山の神田	第36回 日本バイオマテリアル学会大会 (船堀)	2014. 11	国内
PVC製血液バッグに適用可能な 新規可塑剤NJC-NPの毒性評価・ ポスター発表	蓜福野藤山熊井森高河伊柚宮鄭川新吉島井村澤﨑田上川橋上間場崎 雄久見田由千祐彩佳秀 朋美強間俊謙雄久伸田千祐彩佳秀 開美強和志即康一一美吾緑二恵介乃世文薫美和志郎	日本薬学会第135年会 (神戸)	2015. 3	国内

## 2. 学会誌・雑誌等における論文掲載

掲載した論文(発表題目)	発表者氏名	発表した場所 (学会誌・雑誌等名)	発表した時期	国内・外の別
Characterization of alternative plasticizers in polyvinyl chloride sheets for blood containers	Haishima Y, Kawakami T, Fukui C, Tanoue A, Yuba T, Ozono S, Kumada H, Inoue K, Morikawa T, Takahashi M, Fujisawa A, Yamasaki K, Nomura Y, Isama K, Chung U, Ogawa K, Niimi S, Yoshida M.	J. Vinyl Add. Technol. (in press)	2015	国外

Ⅲ. 研究成果の刊行物・別刷

本研究の成果として、以下の論文が平成27年2月16日付けでJ. Vinyl Add. Technol. に採択されたが、平成27年3月現在、発刊に向けた準備中であり、別冊が届いていない。なお、採択通知を次ページに添付した。

Haishima Y, Kawakami T, Fukui C, Tanoue A, Yuba T, Ozono S, Kumada H, Inoue K, Morikawa T, Takahashi M, Fujisawa A, Yamasaki K, Nomura Y, Isama K, Chung U, Ogawa K, Niimi S, Yoshida M.: Characte-rization of alternative plasticizers in polyvinyl chloride sheets for blood containers. *J. Vinyl Add. Technol.*, 2015, in press.

jws1966@ameritech.net 宛先: haishima@nihs.go.jp

Journal of Vinyl and Additive Technology - Decision on Manuscript ID VNL-14-075.R1

16-Feb-2015

Dear Dr. Haishima:

It is a pleasure to accept your manuscript titled "Characterization of alternative plasticizers in polyvinyl chloride sheets for blood containers", in its current form, for publication in the Journal of Vinyl and Additive Technology. The comments of the referee(s) who reviewed your manuscript are included at the bottom of this letter.

Your article cannot be published until the publisher has received the appropriate signed license agreement. Within the next few days the corresponding author will receive an email from Wiley's Author Services system which will ask them to log in and will present them with the appropriate license for completion.

Thank you for your fine contribution.

Sincerely.

Dr. James Summers Editor-in-Chief Journal of Vinyl and Additive Technology jws1966@ameritech.net

Referee(s)' Comments to Author:

Reviewing: 1

#### Comments to the Author

The authors are to be commented for undertaking this important work. I think the paper will be an important contribution to the literature when published; however, I would like to make the strong recommendation to publish the information on the extent to which the plasticizers are released from the polymer and the ability of these plasticizers to inhibit hemolysis separately from the results of the oral toxicity study of DOTH. This would be a much better paper if the authors restricted the scope of the manuscript to address plasticizer release and hemolysis inhibition. The oral toxicity of DOTH is a separate topic that should be the subject of a separate manuscript.

Any statement in the mansucript to support the claim that DOTH is safe to use as a alternative to DEHP as a plasticizer in blood bags should point out that the results of this oral toxicity study may not be directly relevant to the adverse effects produced by DOTH after IV administration (i.e., following infusion of blood products) and that the potency of DOTH may differ depending on the route of administration.

Ideally, it would have been useful to examine other endpoints of RBC function in addition to hemolysis rate. Although hemolysis is an important endpoint, it would be helpful to also examine levels of oxidative stress, ATP, deformability, etc., in RBCs over time following incubation of dilute blood with the polymers containing various levels of the plasticizers. Although this information would have been very useful for the paper, the lack of this information on these endpoints should not serve as a barrier to the publication of the manuscript.

The authors should note that higher release of a new plasticizer compared to one that is commonly used is not necessarily an adverse event if the toxicity of the new plasticizer is less than the exisiting one. Perhaps this can be explained in the manuscript.

The hemolysis study was conducted using blood from on human donor. The sample size of 1 limits the broad interpreation and applicability of the results. RBCs from other donors may be more or less suceptible to hemolysis following incubation than RBCs fro this single donor. This limitation should not prevent the publication of the manuscript, but perhaps some language can be added to underscore the preliminary nature of this work because the hemolysis data represents results from only one individual.

Reviewing: 2

#### Comments to the Author

For the Reader it could be helpful to add an Explanation why the authors changed from DOTP (Dioctyltetrahydrophthalate) as used in Posters to the Name DOTH(Di-octyl-4-cyclohexene-1,2-dicarboxylate; also a short remark like "formerly DOTP was used) could help.

Suggest to consider adding "women in childbearing Age" to the list of Groups where precautions have to be taken.

This is a very nice paper which describes first results with T-die Sheets. I would like to congratulate the authors to These results and would like to motivate them to continue as they proposed in the conclusions, i.e. to Show feasibility with real blood bags.

Please note the Header line of table 2 should be corrected. It should read "DINCH" and not "DINCHI"

Just for precautionary reasons, I would suggest to add a Statement that based on the results of the Pilot blood bag studies, toxicological testing of DOTH on the most likely route of exposure, i.e. the intravenous route is envisaged.

#### Comments:

The comments from the first reviewers have been successfully addressed by the authors. The second set of reviewers have offered a few new suggestions, but I see these more as differing opinions about how data should be interpreted and presented and not a requirement for publication. There was one type-o noted, of the use of DINCHI in Table 2 instead of DINCH. It is a very nice paper and I look forward to reading about future testing of blood containers prepared with your plasticizer blend recommendation.

